

TERMOPRODUKT  
ul.Tkacka 17  
58-260 Bielawa  
tel. +48 (74) 8336 246  
website: www.termoprodukt.co.uk  
e-mail: agatha@termoprodukt.co.uk

# TERMOPRODUKT

MEASURING – CALIBRATION

MEASURING LABORATORY

## CALIBRATION CERTIFICATE

Date of issue:.....

Certificate no:925/23

Page 1/2

**OBJECT OF  
CALIBRATION**

Temperature data logger  
Type: Termio-2  
Serial No: 1530822

**CALIBRATION  
METHOD**

Calibration Procedure 5.4 "Calibration Temperature recorder" edition1 data 4.01.2019

**ENVIRONMENTAL  
CONDITIONS**

Temperature (21,0-22,4)°C  
Humidity (36,6-40,7)%

**DATA OF  
CALIBRATION**

.....

**TRACEABILITY**

Calibration results were referred to reference thermometer maintained in GUM

**CALIBRATION  
RESULTS**

The results have been presented on page 2 of this certificate including uncertainty of measurement

**UNCERTAINTY  
OF MEASUREMENT**

Uncertainty of measurement has been evaluated in compliance with EA 4/02  
The expanded uncertainty assigned corresponds to a coverage probability of 95%  
and the coverage factor  $k=2$

This certificate may be presented or copied as a whole document only

CALIBRATION CERTIFICATE issued by TERMOPRODUKT LABORATORY

Date of issue:.....

Certificate no:925/23

Page 2/2

CALIBRATION RESULTS      Calibration results are the following

Lp	Reference temperature tp °C	Thermometer Termio-2 indication tw °C	Measuring error tw- tp °C	Uncertainty U °C
1	0.00	0.00	0.00	0.1
2	50.00	50.06	0.06	0.1
3	100.00	100.01	0.01	0.1
4	150.00	149.95	-0.05	0.1